

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

| | |
|---------------------------------------|---------------------------------|
| Station: <u>Sheung Wun Yiu (AM1A)</u> | Operator: <u>Chan Wai Shing</u> |
| Cal. Date: <u>30-Jul-12</u> | Next Due Date: <u>30-Sep-12</u> |
| Equipment No.: <u>A-001-53T</u> | Serial No. <u>10216</u> |

| Ambient Condition | | | |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 302 | Pressure, Pa (mmHg) | 752.7 |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|----------|--------------------------------------------------------------------|---------|---------------|----------|
| Serial No: | 843 | Slope, mc | 2.00834 | Intercept, bc | -0.02923 |
| Last Calibration Date: | 5-Nov-11 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 5-Nov-12 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$ | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No. | Orifice | | | HVS Flow Recorder | |
| | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m ³ /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18 | 9.0 | 2.97 | 1.49 | 50.0 | 49.43 |
| 13 | 5.8 | 2.38 | 1.20 | 41.0 | 40.53 |
| 10 | 4.4 | 2.07 | 1.05 | 33.0 | 32.62 |
| 7 | 3.2 | 1.77 | 0.90 | 29.0 | 28.67 |
| 5 | 2.2 | 1.47 | 0.74 | 22.0 | 21.75 |

By Linear Regression of Y on X

Slope, mw = 37.0859 Intercept, bw = -5.2901

Correlation Coefficient* = 0.9916

*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation | |
|------------------------------------------------------------------------------------------|--------------|
| From the TSP Field Calibration Curve, take Qstd = 1.30m ³ /min | |
| From the Regression Equation, the "Y" value according to | |
| $mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$ | |
| Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)] ^{1/2} = | <u>43.42</u> |

Remarks: _____

QC Reviewer: K. H. SHEK Signature: Mike Date: 31. Jul. 12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

| | |
|---------------------------------------|---------------------------------|
| Station: <u>Sheung Wun Yiu (AM1A)</u> | Operator: <u>Gary Choi</u> |
| Cal. Date: <u>28-Sep-12</u> | Next Due Date: <u>28-Nov-12</u> |
| Equipment No.: <u>A-001-53T</u> | Serial No. <u>10216</u> |

| Ambient Condition | | | |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 301 | Pressure, Pa (mmHg) | 755.7 |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|----------|--------------------------------------------------------------------|---------|---------------|----------|
| Serial No: | 843 | Slope, mc | 2.00834 | Intercept, bc | -0.02923 |
| Last Calibration Date: | 5-Nov-11 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 5-Nov-12 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$ | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No. | Orifice | | | HVS Flow Recorder | |
| | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m ³ /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18 | 9.2 | 3.01 | 1.51 | 52.0 | 51.59 |
| 13 | 5.9 | 2.41 | 1.21 | 41.0 | 40.68 |
| 10 | 4.5 | 2.10 | 1.06 | 33.0 | 32.74 |
| 7 | 3.3 | 1.80 | 0.91 | 28.0 | 27.78 |
| 5 | 2.1 | 1.44 | 0.73 | 21.0 | 20.84 |

By Linear Regression of Y on X

Slope, mw = 39.7683

Intercept, bw = -8.4828

Correlation Coefficient* = 0.9967

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m³/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)]^{1/2} =

43.56

Remarks: _____

QC Reviewer: WS CHAN

Signature: RA

Date: 28/9/12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

Station: Shan Tong New Village (AM2) Operator: Shum Kam Yuen
 Cal. Date: 03-Jul-12 Next Due Date: 03-Sep-12
 Equipment No.: A-001-29T Serial No.: 10202

| Ambient Condition | | | |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 306 | Pressure, Pa (mmHg) | 752.7 |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|--------------------------------------------------------------------|---------|---------------|----------|
| Serial No: | 843 | Slope, mc | 2.00834 | Intercept, bc | -0.02923 |
| Last Calibration Date: | 15-Nov-11 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 15-Nov-12 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$ | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No. | Orifice | | | HVS Flow Recorder | |
| | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m ³ /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18 | 9.6 | 3.04 | 1.53 | 48.0 | 47.14 |
| 13 | 7.0 | 2.60 | 1.31 | 42.0 | 41.25 |
| 10 | 5.1 | 2.22 | 1.12 | 34.0 | 33.39 |
| 7 | 4.0 | 1.96 | 0.99 | 28.0 | 27.50 |
| 5 | 2.6 | 1.58 | 0.80 | 22.0 | 21.61 |

By Linear Regression of Y on X

Slope, mw = 36.3998

Intercept, bw = -7.7015

Correlation Coefficient* = 0.9918

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m³/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)]^{1/2} = 40.34

Remarks: _____

QC Reviewer: 

Signature: 

Date: 5-Jul-12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

Station Shan Tong New Village (AM2) Operator: Shum Kam Yuen
 Cal. Date: 3-Sep-12 Next Due Date: 3-Nov-12
 Equipment No.: A-001-29T Serial No. 10202

| Ambient Condition | | | |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 302 | Pressure, Pa (mmHg) | 755.1 |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|---------------------------------------------------------------------|---------|---------------|----------|
| Serial No: | 843 | Slope, mc | 2.00834 | Intercept, bc | -0.02923 |
| Last Calibration Date: | 15-Nov-11 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 15-Nov-12 | $Qstd = \{ [DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc \} / mc$ | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No. | Orifice | | | HVS Flow Recorder | |
| | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m ³ /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18 | 9.5 | 3.05 | 1.53 | 48.0 | 47.53 |
| 13 | 7.0 | 2.62 | 1.32 | 41.0 | 40.60 |
| 10 | 5.2 | 2.26 | 1.14 | 34.0 | 33.67 |
| 7 | 4.0 | 1.98 | 1.00 | 29.0 | 28.71 |
| 5 | 2.5 | 1.57 | 0.79 | 23.0 | 22.77 |

By Linear Regression of Y on X

Slope, mw = 34.0835 Intercept, bw = -4.7902

Correlation Coefficient* = 0.9976

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m³/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)]^{1/2} = 39.91

Remarks:

QC Reviewer: WS CHAN

Signature: RA

Date: 4/9/12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

Station Riverain Bayside (AM3) Operator: Shum Kam Yuen
 Cal. Date: 03-Jul-12 Next Due Date: 03-Sep-12
 Equipment No.: A-001-69T Serial No. 716

| Ambient Condition | | | |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 306 | Pressure, Pa (mmHg) | 752.7 |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|--------------------------------------------------------------------|---------|---------------|----------|
| Serial No: | 843 | Slope, mc | 2.00834 | Intercept, bc | -0.02923 |
| Last Calibration Date: | 15-Nov-11 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 15-Nov-12 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$ | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No. | Orifice | | | HVS Flow Recorder | |
| | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m ³ /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18 | 8.4 | 2.85 | 1.43 | 44.0 | 43.21 |
| 13 | 7.6 | 2.71 | 1.36 | 42.0 | 41.25 |
| 10 | 6.1 | 2.43 | 1.22 | 38.0 | 37.32 |
| 7 | 4.8 | 2.15 | 1.09 | 32.0 | 31.43 |
| 5 | 3.0 | 1.70 | 0.86 | 24.0 | 23.57 |

By Linear Regression of Y on X

Slope, mw = 34.9045 Intercept, bw = -6.2805

Correlation Coefficient* = 0.9953

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m³/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC = $(mw \times Qstd + bw) \times [(760 / Pa) \times (Ta / 298)]^{1/2} =$ 39.81

Remarks: _____

QC Reviewer: [Signature]

Signature: [Signature]

Date: 5 Jul 12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

| | |
|----------------------------------------|--------------------------------|
| Station: <u>Riverain Bayside (AM3)</u> | Operator: <u>Shum Kam Yuen</u> |
| Cal. Date: <u>3-Sep-12</u> | Next Due Date: <u>3-Nov-12</u> |
| Equipment No.: <u>A-001-69T</u> | Serial No. <u>716</u> |

| Ambient Condition | | | |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 302 | Pressure, Pa (mmHg) | 755.1 |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|--------------------------------------------------------------------|---------|---------------|----------|
| Serial No: | 843 | Slope, mc | 2.00834 | Intercept, bc | -0.02923 |
| Last Calibration Date: | 15-Nov-11 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 15-Nov-12 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$ | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No. | Orifice | | | HVS Flow Recorder | |
| | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m ³ /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18 | 8.6 | 2.90 | 1.46 | 45.0 | 44.56 |
| 13 | 7.8 | 2.77 | 1.39 | 42.0 | 41.59 |
| 10 | 6.2 | 2.47 | 1.24 | 36.0 | 35.65 |
| 7 | 4.9 | 2.19 | 1.11 | 33.0 | 32.67 |
| 5 | 3.0 | 1.71 | 0.87 | 24.0 | 23.76 |

By Linear Regression of Y on X

Slope, mw = 34.2608

Intercept, bw = -5.9363

Correlation Coefficient* = 0.9935

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m³/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)]^{1/2} =

38.99

Remarks: _____

QC Reviewer: WS CHAN

Signature: RA

Date: 4/9/12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

| | |
|-------------------------------------------|--------------------------|
| Station: 168 Shek Kwu Lung Village (AM4A) | Operator: Chan Wai Shing |
| Cal. Date: 30-Jul-12 | Next Due Date: 30-Sep-12 |
| Equipment No.: A-001-70T | Serial No. 10273 |

| Ambient Condition | | | |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 302 | Pressure, Pa (mmHg) | 752.7 |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|--------------------------------------------------------------------|---------|---------------|----------|
| Serial No: | 843 | Slope, mc | 2.00834 | Intercept, bc | -0.02923 |
| Last Calibration Date: | 15-Nov-11 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | | | |
| Next Calibration Date: | 15-Nov-12 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$ | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No. | Orifice | | | HVS Flow Recorder | |
| | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m ³ /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18 | 9.9 | 3.11 | 1.56 | 50.0 | 49.43 |
| 13 | 7.4 | 2.69 | 1.35 | 44.0 | 43.50 |
| 10 | 5.2 | 2.25 | 1.14 | 36.0 | 35.59 |
| 7 | 3.8 | 1.93 | 0.97 | 31.0 | 30.65 |
| 5 | 2.2 | 1.47 | 0.74 | 24.0 | 23.73 |

By Linear Regression of Y on X

Slope, mw = 31.8601 Intercept, bw = -0.2064

Correlation Coefficient* = 0.9985

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m³/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)]^{1/2} = 41.69

Remarks:

QC Reviewer: K. M. SHEK

Signature: Mike

Date: 31 Jul 12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

| | |
|-------------------------------------------|--------------------------|
| Station: 168 Shek Kwu Lung Village (AM4A) | Operator: Gary Choi |
| Cal. Date: 28-Sep-12 | Next Due Date: 28-Nov-12 |
| Equipment No.: A-001-70T | Serial No.: 10273 |

| Ambient Condition | | | |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 301 | Pressure, Pa (mmHg) | 755.7 |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|-----------------------------------------------------------------------------------------------------------------------------------------|---------|---------------|----------|
| Serial No: | 843 | Slope, mc | 2.00834 | Intercept, bc | -0.02923 |
| Last Calibration Date: | 15-Nov-11 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$ | | | |
| Next Calibration Date: | 15-Nov-12 | | | | |

| Calibration of TSP Sampler | | | | | |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No. | Orifice | | | HVS Flow Recorder | |
| | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m ³ /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18 | 9.6 | 3.07 | 1.55 | 50.0 | 49.61 |
| 13 | 7.4 | 2.70 | 1.36 | 45.0 | 44.65 |
| 10 | 5.3 | 2.28 | 1.15 | 36.0 | 35.72 |
| 7 | 3.6 | 1.88 | 0.95 | 30.0 | 29.77 |
| 5 | 2.2 | 1.47 | 0.75 | 22.0 | 21.83 |

By Linear Regression of Y on X

Slope, mw = 35.2097 Intercept, bw = -4.2114

Correlation Coefficient* = 0.9958

*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <p>From the TSP Field Calibration Curve, take Qstd = 1.30m³/min</p> <p>From the Regression Equation, the "Y" value according to</p> $mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$ <p>Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)]^{1/2} = <u>41.89</u></p> |

Remarks: _____

QC Reviewer: WS CHAN

Signature: RL

Date: 28/9/12



TISCH ENVIRONMENTAL, INC.
145 SOUTH MIAMI AVE.
VILLAGE OF CLEVELAND, OH 45002
513.467.9000
877.263.7610 TOLL FREE
513.467.9009 FAX
WWW.TISCH-ENV.COM

AIR POLLUTION MONITORING EQUIPMENT

ORIFICE TRANSFER STANDARD CERTIFICATION WORKSHEET TE-5025A

Date - Nov 15, 2011 Rootsmeter S/N 0438320 Ta (K) - 294
Operator Tisch Orifice I.D. - 0843 Pa (mm) - 748.03

| PLATE OR Run # | VOLUME START (m3) | VOLUME STOP (m3) | DIFF VOLUME (m3) | DIFF TIME (min) | METER DIFF Hg (mm) | ORFICE DIFF H2O (in.) |
|----------------------|-------------------------|------------------------|------------------------|-----------------------|-----------------------------|--------------------------------|
| 1 | NA | NA | 1.00 | 1.3810 | 3.2 | 2.00 |
| 2 | NA | NA | 1.00 | 0.9810 | 6.4 | 4.00 |
| 3 | NA | NA | 1.00 | 0.8760 | 7.8 | 5.00 |
| 4 | NA | NA | 1.00 | 0.8370 | 8.8 | 5.50 |
| 5 | NA | NA | 1.00 | 0.6890 | 12.7 | 8.00 |

DATA TABULATION

| Vstd | (x axis) Qstd | (y axis) | | Va | (x axis) Qa | (y axis) |
|-------------------------------------|------------------|----------|--|---------------------------|----------------|----------|
| 0.9934 | 0.7193 | 1.4125 | | 0.9957 | 0.7210 | 0.8866 |
| 0.9891 | 1.0083 | 1.9976 | | 0.9915 | 1.0107 | 1.2538 |
| 0.9871 | 1.1269 | 2.2334 | | 0.9895 | 1.1295 | 1.4018 |
| 0.9859 | 1.1779 | 2.3424 | | 0.9882 | 1.1807 | 1.4703 |
| 0.9807 | 1.4233 | 2.8251 | | 0.9830 | 1.4267 | 1.7732 |
| Qstd slope (m) = 2.00834 | | | | Qa slope (m) = 1.25759 | | |
| intercept (b) = -0.02923 | | | | intercept (b) = -0.01835 | | |
| coefficient (r) = 0.99994 | | | | coefficient (r) = 0.99994 | | |
| y axis = SQRT[H2O(Pa/760) (298/Ta)] | | | | y axis = SQRT[H2O(Ta/Pa)] | | |

CALCULATIONS

Vstd = Diff. Vol [(Pa-Diff. Hg)/760] (298/Ta)
Qstd = Vstd/Time

Va = Diff Vol [(Pa-Diff Hg)/Pa]
Qa = Va/Time

For subsequent flow rate calculations:

Qstd = 1/m{ [SQRT(H2O(Pa/760) (298/Ta))] - b}
Qa = 1/m{ [SQRT H2O(Ta/Pa)] - b}

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
Manufacturer/Brand: SIBATA
Model No.: LD-3
Equipment No.: A.005.07a
Sensitivity Adjustment Scale Setting: 557 CPM

Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
Venue: Cyberport (Pui Ying Secondary School)
Model No.: Series 1400AB
Serial No: Control: 140AB219899803
Sensor: 1200C143659803 K₀: 12500
Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 557 CPM
Sensitivity Adjustment Scale Setting (After Calibration): 557 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|----------------------|-------------|--------------------------------------------------------------|-----------------------------|-----------------------------------------|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 02-06-12 | 13:30 - 14:30 | 27.9 | 63 | 0.04070 | 1628 | 27.13 |
| 2 | 02-06-12 | 14:30 - 15:30 | 27.9 | 63 | 0.04167 | 1669 | 27.82 |
| 3 | 02-06-12 | 15:30 - 16:30 | 28.2 | 64 | 0.04283 | 1713 | 28.55 |
| 4 | 02-06-12 | 16:30 - 17:30 | 28.1 | 63 | 0.04146 | 1655 | 27.58 |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
2. Total Count was logged by Laser Dust Monitor
3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
Correlation coefficient: 0.9951

Validity of Calibration Record: 1 June 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 4 June 2012

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.08a
 Sensitivity Adjustment Scale Setting: 702 CPM
 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 702 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 702 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|-------------------|-------------|--------------------------------------------------------------|--------------------------|-----------------------------------------|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 02-07-12 | 13:30 - 14:30 | 28.9 | 73 | 0.04127 | 1545 | 25.75 |
| 2 | 02-07-12 | 14:30 - 15:30 | 29.0 | 73 | 0.04163 | 1566 | 26.10 |
| 3 | 02-07-12 | 15:30 - 16:30 | 29.0 | 73 | 0.04334 | 1630 | 27.17 |
| 4 | 02-07-12 | 16:30 - 17:30 | 29.1 | 74 | 0.04426 | 1665 | 27.74 |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0016
 Correlation coefficient: 0.9952

Validity of Calibration Record: 1 July 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 3 July 2012

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.09a
 Sensitivity Adjustment Scale Setting: 797 CPM
 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No.: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 797 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 797 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|----------------------|-------------|--------------------------------------------------------------|-----------------------------|-----------------------------------------|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 02-06-12 | 13:30 - 14:30 | 27.9 | 63 | 0.04070 | 1626 | 27.10 |
| 2 | 02-06-12 | 14:30 - 15:30 | 27.9 | 63 | 0.04167 | 1667 | 27.78 |
| 3 | 02-06-12 | 15:30 - 16:30 | 28.2 | 64 | 0.04283 | 1708 | 28.47 |
| 4 | 02-06-12 | 16:30 - 17:30 | 28.1 | 63 | 0.04146 | 1659 | 27.65 |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9949

Validity of Calibration Record: 1 June 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 4 June 2012

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
Manufacturer/Brand: SIBATA
Model No.: LD-3
Equipment No.: A.005.10a
Sensitivity Adjustment Scale Setting: 753 CPM
Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
Venue: Cyberport (Pui Ying Secondary School)
Model No.: Series 1400AB
Serial No: Control: 140AB219899803
Sensor: 1200C143659803 K₀: 12500
Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 753 CPM
Sensitivity Adjustment Scale Setting (After Calibration): 753 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|----------------------|-------------|--------------------------------------------------------------|-----------------------------|-----------------------------------------|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 02-06-12 | 12:45 - 13:45 | 27.9 | 63 | 0.04041 | 1613 | 26.88 |
| 2 | 02-06-12 | 13:45 - 14:45 | 27.9 | 63 | 0.04085 | 1631 | 27.18 |
| 3 | 02-06-12 | 14:45 - 15:45 | 27.9 | 63 | 0.04154 | 1663 | 27.72 |
| 4 | 02-06-12 | 15:45 - 16:45 | 28.1 | 64 | 0.04272 | 1711 | 28.52 |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
2. Total Count was logged by Laser Dust Monitor
3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
Correlation coefficient: 0.9939

Validity of Calibration Record: 1 June 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 4 June 2012

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.11a
 Sensitivity Adjustment Scale Setting: 799 CPM
 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 799 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 799 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|-------------------|-------------|--------------------------------------------------------------|--------------------------|-----------------------------------------|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 02-07-12 | 13:45 - 14:45 | 29.0 | 73 | 0.04152 | 1659 | 27.65 |
| 2 | 02-07-12 | 14:45 - 15:45 | 29.0 | 73 | 0.04194 | 1670 | 27.83 |
| 3 | 02-07-12 | 15:45 - 16:45 | 29.1 | 74 | 0.04318 | 1725 | 28.75 |
| 4 | 02-07-12 | 16:45 - 17:45 | 29.1 | 74 | 0.04443 | 1780 | 29.67 |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9928

Validity of Calibration Record: 1 July 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 3 July 2012

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
Manufacturer/Brand: SIBATA
Model No.: LD-3B
Equipment No.: A.005.14a
Sensitivity Adjustment Scale Setting: 786 CPM

Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
Venue: Cyberport (Pui Ying Secondary School)
Model No.: Series 1400AB
Serial No: Control: 140AB219899803
Sensor: 1200C143659803 K_o: 12500
Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 786 CPM
Sensitivity Adjustment Scale Setting (After Calibration): 786 CPM

| Hour | Date (dd-mm-yy) | Time | Ambient Condition | | Concentration ¹ (mg/m ³) Y-axis | Total Count ² | Count/ Minute ³ X-axis |
|------|--------------------|---------------|----------------------|-------------|--------------------------------------------------------------|-----------------------------|-----------------------------------------|
| | | | Temp (°C) | R.H. (%) | | | |
| 1 | 02-06-12 | 13:15 - 14:15 | 27.9 | 63 | 0.04073 | 1746 | 29.10 |
| 2 | 02-06-12 | 14:15 - 15:15 | 27.9 | 63 | 0.04154 | 1778 | 29.63 |
| 3 | 02-06-12 | 15:15 - 16:15 | 28.1 | 64 | 0.04269 | 1830 | 30.50 |
| 4 | 02-06-12 | 16:15 - 17:15 | 28.1 | 64 | 0.04136 | 1769 | 29.48 |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
2. Total Count was logged by Laser Dust Monitor
3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0014
Correlation coefficient: 0.9963

Validity of Calibration Record: 1 June 2013

Remarks:

QC Reviewer: YW Fung Signature:  Date: 4 June 2012



CERTIFICATE OF CALIBRATION

Certificate No.: 11CA1221 01-01

Page 1 of 2

Item tested

| | | | |
|-----------------------|----------------------------|----------------|----------------|
| Description: | Sound Level Meter (Type 1) | Microphone | Preamp |
| Manufacturer: | Rion Co., Ltd. | Rion Co., Ltd. | Rion Co., Ltd. |
| Type/Model No.: | NL-31 | UC-53A | NH-21 |
| Serial/Equipment No.: | 00320534 / N.007.02A | 90526 | 03581 |
| Adaptors used: | - | - | - |

Item submitted by

Customer Name: AECOM ASIA CO., LTD.
Address of Customer: -
Request No.: -
Date of receipt: 21-Dec-2011

Date of test: 23-Dec-2011

Reference equipment used in the calibration

| Description: | Model: | Serial No. | Expiry Date: | Traceable to: |
|---------------------------------|----------|------------|--------------|---------------|
| Multi function sound calibrator | B&K 4226 | 2288444 | 09-May-2012 | CIGISMEC |
| Signal generator | DS 360 | 33873 | 30-May-2012 | CEPREI |
| Signal generator | DS 360 | 61227 | 30-May-2012 | CEPREI |

Ambient conditions

Temperature: (22 ± 1) °C
Relative humidity: (60 ± 10) %
Air pressure: (1000 ± 5) hPa

Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of ±20%.
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responsiveness of the Sound Level Meter.

Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.

Actual Measurement data are documented on worksheets.

Approved Signatory:


Huang Jian Min/Feng Jun Qi

Date: 16-Jan-2012

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



CERTIFICATE OF CALIBRATION

Certificate No.: 12CA0321 01-02

Page 1 of 2

Item tested

| | | |
|-----------------------|----------------------------|------------|
| Description: | Sound Level Meter (Type 1) | Microphone |
| Manufacturer: | B & K | B & K |
| Type/Model No.: | 2250-L | 4950 |
| Serial/Equipment No.: | 2681366 / N.011.01 | 2665582 |
| Adaptors used: | - | - |

Item submitted by

Customer Name: AECOM ASIA CO LTD
Address of Customer: -
Request No.: -
Date of receipt: 21-Mar-2012

Date of test: 23-Mar-2012

Reference equipment used in the calibration

| Description: | Model: | Serial No. | Expiry Date: | Traceable to: |
|---------------------------------|----------|------------|--------------|---------------|
| Multi function sound calibrator | B&K 4226 | 2288444 | 09-May-2012 | CIGISMEC |
| Signal generator | DS 360 | 33873 | 30-May-2012 | CEPREI |
| Signal generator | DS 360 | 61227 | 30-May-2012 | CEPREI |

Ambient conditions

Temperature: (22 ± 1) °C
Relative humidity: (60 ± 10) %
Air pressure: (1005 ± 5) hPa

Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of ±20%.
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responsiveness of the Sound Level Meter.


Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.

Actual Measurement data are documented on worksheets.

Approved Signatory:


Huang Jian Min Feng Jun Qi

Date: 23-Mar-2012

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



CERTIFICATE OF CALIBRATION

Certificate No.: 11CA1221 01-02

Page: 1 of 2

Item tested

Description: Acoustical Calibrator (Class 1)
Manufacturer: Rion Co., Ltd.
Type/Model No.: NC-73
Serial/Equipment No.: 10307216 / N.004.06
Adaptors used: -

Item submitted by

Customer: AECOM ASIA CO., LTD.
Address of Customer: -
Request No.: -
Date of receipt: 21-Dec-2011

Date of test: 16-Jan-2012

Reference equipment used in the calibration

| Description: | Model: | Serial No. | Expiry Date: | Traceable to: |
|-------------------------|----------|------------|--------------|---------------|
| Lab standard microphone | B&K 4180 | 2341427 | 18-May-2012 | SCL |
| Preamplifier | B&K 2673 | 2239857 | 14-Dec-2011 | CEPREI |
| Measuring amplifier | B&K 2610 | 2346941 | 15-Dec-2011 | CEPREI |
| Signal generator | DS 360 | 61227 | 30-May-2012 | CEPREI |
| Digital multi-meter | 34401A | US36087050 | 09-Dec-2011 | CEPREI |
| Audio analyzer | 8903B | GB41300350 | 27-May-2012 | CEPREI |
| Universal counter | 53132A | MY40003662 | 30-May-2012 | CEPREI |

Ambient conditions

Temperature: 22 ± 1 °C
Relative humidity: 65 ± 5 %
Air pressure: 1005 ± 5 hPa

Test specifications

- 1, The Sound Calibrator has been calibrated in accordance with the requirements as specified in IEC 60942 1997 Annex B and the lab calibration procedure SMTP004-CA-156.
- 2, The calibrator was tested with its axis vertical facing downwards at the specific frequency using insert voltage technique.
- 3, The results are rounded to the nearest 0.01 dB and 0.1 Hz and have not been corrected for variations from a reference pressure of 1013.25 hectoPascals as the maker's information indicates that the instrument is insensitive to pressure changes.

Test results

This is to certify that the sound calibrator conforms to the requirements of annex B of IEC 60942: 1997 for the conditions under which the test was performed. This does not imply that the sound calibrator meets IEC 60942 under any other conditions.

Details of the performed measurements are presented on page 2 of this certificate.

Approved Signatory:

Huang Jian Min Feng Jun Qi

Date: 16-Jan-2012

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



CERTIFICATE OF CALIBRATION

Certificate No.: 12CA0817 01

Page: 1 of 2

Item tested

Description: Acoustical Calibrator (Class 1)
Manufacturer: Rion Co., Ltd.
Type/Model No.: NC-73
Serial/Equipment No.: 10307223 / N.004.08
Adaptors used: -

Item submitted by

Customer: AECOM ASIA CO., LTD.
Address of Customer: -
Request No.: -
Date of receipt: 17-Aug-2012

Date of test: 17-Aug-2012

Reference equipment used in the calibration

| Description: | Model: | Serial No. | Expiry Date: | Traceable to: |
|-------------------------|----------|------------|--------------|---------------|
| Lab standard microphone | B&K 4180 | 2412857 | 29-May-2013 | SCL |
| Preamplifier | B&K 2673 | 2239857 | 05-Jan-2013 | CEPREI |
| Measuring amplifier | B&K 2610 | 2346941 | 29-Dec-2012 | CEPREI |
| Signal generator | DS 360 | 61227 | 29-May-2013 | CEPREI |
| Digital multi-meter | 34401A | US36087050 | 16-Dec-2012 | CEPREI |
| Audio analyzer | 8903B | GB41300350 | 29-May-2013 | CEPREI |
| Universal counter | 53132A | MY40003662 | 29-May-2013 | CEPREI |

Ambient conditions

Temperature: 22 ± 1 °C
Relative humidity: 60 ± 10 %
Air pressure: 995 ± 5 hPa

Test specifications

- 1, The Sound Calibrator has been calibrated in accordance with the requirements as specified in IEC 60942 1997 Annex B and the lab calibration procedure SMTP004-CA-156.
- 2, The calibrator was tested with its axis vertical facing downwards at the specific frequency using insert voltage technique.
- 3, The results are rounded to the nearest 0.01 dB and 0.1 Hz and have not been corrected for variations from a reference pressure of 1013.25 hectoPascals as the maker's information indicates that the instrument is insensitive to pressure changes.

Test results

Details of the performed measurements are presented on page 2 of this certificate.

Approved Signatory:

Huang Jian Min/Feng Jun Qi

Date: 17-Aug-2012

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.